

Certificate of Analysis

Product Name:	AcroMetrix EV Negative Control, CE-IVD
Catalog Number:	960499
Lot Number:	237219
Expiration Date (YYYY-MM-DD):	2023-08-31
Viral Source:	N/A
Diluent:	OptiMatrix Clear
Preservative:	ProClin 300
Fill Volume:	0.3 mL
Storage:	-70°C or below

Summary of Quality Control Release Testing Results

Extraction Method	Assay	Results
Qiagen MinElute Virus Spin Kit	In-House Enterovirus Real-Time Assay	Pass

Test	Acceptance Criteria	Results
Spreadplate Bioburden	≤ 10 CFU/mL	0 CFU/mL

Precautions

Observe universal precautions for prevention of transmission of infectious agents when handling these materials.

Intended Use

This product is for In Vitro Diagnostic use. This product is CE-marked.

CERTIFICATION

The product described above meets all Quality Control performance specification:

Signature: *Grace Hayes*

Date: 03 SEP 2021