ThermoFisher SCIENTIFIC

Certificate of Analysis

Product No.: 088602

Product: Potassium hydroxide, Acculute Standard Volumetric Solution, Final Concentration 0.5N

Lot No.: R16J002

Appearance Clear liquid

Assay 0.5N±0.001N

Certification and Traceability: Thermo Fisher Scientific is ISO 9001:2015 certified. This CRM was manufactured and certified by a Thermo Fisher Scientific supplier under ISO 9001, ISO/IEC 17025, and ISO 17034 Quality Management System.

Expiration date: November 16, 2024

Order our products online **thermofisher.com/chemicals**

This document has been electronically generated and does not require a signature.

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.