ThermoFisher SCIENTIFIC

Certificate of Analysis

Product No.: 042131

Product: ORP test solution, 200-275 mV

Lot No.: R04H007

Mean value 233 mV

Method of standardization:

This quality product is standardized using a ORP meter and ORP electrode periodically checked. All primary standard solutions used are prepared with deionized water for analytical use. Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95 % level of confidence.

Expiration date: May 4, 2026

Order our products online **thermofisher.com/chemicals**

This document has been electronically generated and does not require a signature.

Products are processed under ISO 9001:2015 quality management systems and samples are tested for conformance to the noted specifications. Certain data may have been supplied by third parties. We disclaim the implied warranties of merchantability and fitness for a particular purpose, and the accuracy of third party data or information associated with the product. Products are for research and development use only. Products are not for direct administration to humans or animals. It is the responsibility of the final formulator or end user to determine suitability, and to qualify and/or validate each product for its intended use.