

Certificate of analysis

Product No.: 47387

Multi-Element ICH/USP Parenteral Combined Standard Solution G

Matrix: 15% HCl

Lot No.: 8137939

Expiry Date: April 30, 2020

Intended Use: This solution is intended for use as a certified reference material or quality control standard for inductively coupled plasma optical emission spectroscopy (ICP-OES), inductively coupled plasma mass spectrometry (ICP-MS). It is designed to meet the needs of the analysis of elemental impurities in pharmaceutical drug products based on guidelines set in ICH Q3D, chapters USP <232> and USP <233>.

Element	Certified Concentration & Uncertainty (w/v)	Certified Concentration & Uncertainty (w/w)
Au	100.0 ± 0.50 µg/mL	97.84 ± 0.49 µg/g
Ir	10.00 ± 0.05 µg/mL	9.78 ± 0.05 µg/g
Os	10.00 ± 0.05 µg/mL	9.79 ± 0.05 µg/g
Pd	10.00 ± 0.05 µg/mL	9.78 ± 0.05 µg/g
Pt	10.00 ± 0.05 µg/mL	9.79 ± 0.05 µg/g
Rh	10.00 ± 0.05 µg/mL	9.78 ± 0.05 µg/g
Ru	10.00 ± 0.05 µg/mL	9.78 ± 0.05 µg/g

Certification & Traceability: This CRM was manufactured and certified under an ISO 9001, ISO/IEC 17025, and ISO Guide 34 quality management system. This CRM was prepared to the certified concentrations shown above by gravimetric methods using single-element concentrates that were certified using the "High Performance ICP-OES" protocol developed by NIST and are directly traceable to the NIST SRMs listed below. This solution was stabilized using high purity hydrochloric acid (HCl) and diluted with filtered (0.22µm), 18 M-ohm deionized water. The balances used in the preparation of this CRM are calibrated regularly with traceability to NIST. All volumetric dilutions are performed in Class A calibrated glassware. The certified concentrations were determined based upon gravimetric procedures. Secondary verification of the certified concentrations was performed using ICP-OES that was calibrated and/or referenced against NIST SRMs: 3121, 3138, 3140, and 3144. The uncertainty associated with each certified concentration represents the expanded uncertainty at the 95% confidence level using a coverage factor of k=2.

Instructions for Use: We recommend that the solution be thoroughly mixed by repeated shaking or swirling of the bottle immediately prior to use. To achieve the highest accuracy the analyst should: (1) use only pre-cleaned containers and transferware, (2) not pipette directly from the CRM's original container, (3) use a minimum sub-sample size of 500µL, (4) make dilutions using calibrated balances or certified volumetric class A flasks and pipettes, (5) dilute with the same matrix as the original CRM, and (6) never pour used product back into the original container. The solution should be kept tightly capped and stored under normal laboratory conditions. Do not freeze, heat, or expose to direct sunlight. Minimize exposure to moisture or high humidity.

Period of Validity: Alfa Aesar guarantees the accuracy of this Specpure® solution until the expiry date shown above, provided the instructions for use are followed. During the period of validity, the purchaser will be notified if this product is recalled due to any significant changes in the stability of the solution.

10/5/2018

Certification Date

This document has been electronically generated and does not require a signature.

Page 1 of 2

Certificate of analysis

Hazard Information: Refer to the Safety Data Sheet (SDS).

Homogeneity: This solution was determined to be homogeneous by procedures consistent with the requirements of ISO Guide 34 and ISO Guide 35. Replicate samples of the finished solution were analyzed to confirm its homogeneity, in accordance with QSP 6-13 Assessment of Homogeneity and Stability. To ensure homogeneity, users should not take a smaller sub-sample than specified in the Instructions for Use, as doing so will invalidate the certified values and uncertainties.

Further Information: Please contact Alfa Aesar for further information about this CRM.

Quality Certifications: This CRM was prepared under a quality management system that is:

- registered to ISO 9001 – Quality Management Systems – Requirements (TUV NORD Cert. No. 44 100 16560231)
- accredited to ISO Guide 34 – General Requirements for the Competence of Reference Material Producers (A2LA Cert. No. 2848.02)
 - ISO Guide 34 references additional requirements specified in ISO Guide 31 and ISO Guide 35
- accredited ISO/IEC 17025 – General Requirements for the Competence of Testing and Calibration Laboratories (A2LA Cert. No. 2848.01)

This document has been electronically generated and does not require a signature.

Page 2 of 2

Order our products online www.alfa.com

ThermoFisher
S C I E N T I F I C