

Certificate of Analysis

PLURONIC F-68 SOLN

Origin : UK

Catalogue Number : 780-00151

Lot Number : 1907359

Expiration Date : 2018-05

TEST	SPECIFICATION	RESULT	UNITS
q01 Documentation		Satisfactory	
srt00063 Sample Retained?	Yes	Yes	
srt00087 Pluronic F68 (Poloxamer 188) Performance Assay		Acceptable	
srt00103 Pluronic F68 Performance Assay – Test Result		3.28	x10^6 ViableCells/mL
srt00104 Pluronic F68 Performance Assay – Test Viability		97	%
srt00105 Pluronic F68 Performance Assay – Control Result		2.94	x10^6 ViableCells/mL
srt00106 Pluronic F68 Performance Assay – Control Viability		95	%

M. M'Fadyen

Quality System Department

Date : 07-Aug-2017