



## Certificate of Quality

**Product Name:** Chocolate Agar w/Bacitracin 10/PK  
**Lot Number:** 145424

**Product Number:** R01304  
**Expiration Date:** 2012-04-03  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Haemophilus influenzae	10211	Good Growth
Staphylococcus aureus	25923	Marked Inhibition
Streptococcus sanguis	10556	No Growth

**Physical Characteristics:**

Appearance: Brown, Opaque Agar  
pH @ 20-25°C: 7.2 +/- 0.2

**Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell". The signature is written in a cursive, flowing style.

Quality Assurance Supervisor